

# Rapidos® | Short-Term Radon Test for Dwellings

## INSTRUCTION FOR SHORT-TERM RADON MEASUREMENT

- 1** Read these instructions in their entirety. The detector(s) should be placed accordingly, with the measurement period being at least 10 days, not exceeding 90 days. The measurement begins when the radon-proof plastic bag is opened. Immediately place device.  
Go to [www.radonline.com](http://www.radonline.com) and use your commission number and password that is provided to log in. Alternatively, scanning the QR code provided with any smart device will take you to the commission and log in automatically. iOS users (11+) can scan the code with their camera. Android devices will need an app, QR Scanner or similar. Fill in measurement information, including specific detector location, detector number (*which is located on the bottom of the detector*) and start date and time. The detector(s) should remain deployed in the same location until they are ready to be returned.
- 2** Upon completion of test, log in to My Pages and provide the measurement end date and time. Confirm your location information and Save. Package up your detector(s) and return immediately to the laboratory address below. Retain the documents that include your commission number and password for your records.
- 3** Upon receipt your detector(s) will be analyzed. Measurement Reports are typically available within 5-10 business days. The report can be emailed to the email address provided, and may be retrievable from the website as well.
- 4**

## RECOMMENDED DETECTOR PLACEMENT

- At least 20" above the floor
- At least 3' from windows or exterior doors
- At least 4" from objects or outside walls
- **Do not place on any mantels or other natural stone**

**Please note!** Some states or organizations may have differing recommendations. If you have a question regarding placement, contact your state radon agency.

- The EPA recommends the test kit be placed in the lowest, lived-in level of the home (i.e. a frequently used basement)
- It should be put in a room that is regularly used such as a living room, den, or bedroom, **but not in a kitchen or bathroom**

## DEPLOYMENT PROTOCOL

Short-term measurements lasting between 10 and 90 days should be made under closed-building conditions. Windows on all levels and external doors should be kept closed (except during normal entry and exit) during the measurement period. In addition, external-internal air exchange systems (other than a furnace) such as high-volume, whole-house and window fans should not be operating. However, attic fans intended to control attic and not whole building temperature or humidity should continue to operate. Combustion or make-up air supplies must not be closed.

## RETURN DETECTOR(S) TO

Radonova, Inc.  
900 Oakmont Lane Suite 207, Westmont IL 60559

## QUESTIONS REGARDING THE MEASUREMENT

Email: [help@radonova.com](mailto:help@radonova.com)  
Telephone: 331.814.2200

DISCLAIMER –Radonova, Inc. makes no warranty of any kind, express or implied, as regard to the use, operation or analysis of any Radonova, Inc. monitor. Radonova, Inc. specifically disclaims implied warranties of merchantability and fitness for a particular purpose. Radonova, Inc. is not responsible for any damage, including consequential damages, to persons or property resulting from the use of the monitor or the resulting data.